



Member of the FM Global Group

FM Approvals
1151 Boston Providence Turnpike
P.O. Box 9102 Norwood, MA 02062 USA
T: 781 762 4300 F: 781-762-9375 www.fmaprovals.com

CERTIFICATE OF COMPLIANCE

HAZARDOUS (CLASSIFIED) LOCATION ELECTRICAL EQUIPMENT

This certificate is issued for the following equipment:

CSI2140. Machinery Health Analyzer

NI / I / 2 / ABCD / T6, Ta = 50°C — D25533; NIFW; IP65

NI / I / 2 / IIC / T6, Ta = 50°C — D25533; NIFW; IP65

Special Conditions of use:

The USB, Ethernet port and Tachometer are not to be used in the hazardous location.

Equipment Ratings:

Nonincendive with Nonincendive Field Wiring Connections for use in Class I, Division 2, Groups A, B, C and D IP65 indoor and outdoor hazardous (classified) locations.

FM Approved for:

Computational Systems Inc.
835 Innovation Drive
Knoxville TN 37932



Member of the FM Global Group

This certifies that the equipment described has been found to comply with the following Approval Standards and other documents:

Class 3611	2004
Class 3600	2011
Class 3810	2005
ISA 60529	2003

Original Project ID: 3048590

Approval Granted: May 22, 2014

Subsequent Revision Reports / Date Approval Amended

Report Number	Date	Report Number	Date
---------------	------	---------------	------

FM Approvals LLC

J.E. Marquedant
Manager, Electrical Systems

22 May 2014

Date